Clinical Neurophysiology 127 (2016) 3498

Contents lists available at ScienceDirect

Clinical Neurophysiology

journal homepage: www.elsevier.com/locate/clinph

Corrigendum

Corrigendum to "Improved multi-stage neonatal seizure detection using a heuristic classifier and a data-driven post-processor" [Clin Neurophysiol 127 (2016) 3014–3024]

A.H. Ansari [a,b,*], P.J. Cherian [c,d], A. Dereymaeker [e], V. Matic [a,b], K. Jansen [e,f], L. De Wispelaere [g], C. Dielman [h], J. Vervisch [e], R.M. Swarte [g], P. Govaert [g,h], G. Naulaers [e], M. De Vos [i], S. Van Huffel [a,b]

[a] Department of Electrical Engineering (ESAT), KU Leuven, Leuven, Belgium
[b] iMinds Medical Information Technology, Leuven, Belgium
[c] Section of Clinical Neurophysiology, Department of Neurology, Erasmus MC, University Medical Center Rotterdam, The Netherlands
[d] Division of Neurology, Department of Medicine, McMaster University, Hamilton, Canada
[e] Department of Development and Regeneration, University Hospitals Leuven, Neonatal Intensive Care Unit, KU Leuven, Leuven, Belgium
[f] Department of Development and Regeneration, University Hospitals Leuven, Child Neurology, KU Leuven, Leuven, Belgium
[g] Section of Neonatology, Department of Pediatrics, Sophia Children's Hospital, Erasmus MC, University Medical Center Rotterdam, The Netherlands
[h] ZNA Koningin Paola Kinderziekenhuis, Antwerp, Belgium
[i] Institute of Biomedical Engineering, Department of Engineering, University of Oxford, Oxford, UK

The authors regret that some text regarding Fig. 3 inadvertently appeared on p. 3018 in the paragraph after equation (3). The text should have read:

To illustrate the difference of MPC between an artifact and an actual seizure, Fig. 4 plots a falsely detected ECG artifact (a, MPC: 0:21) and a truly detected rhythmic seizure (b, MPC: 0:04). In the current work, the MPC of EEG segment and the polygraphic signals ECG, EMG, EOG, and Resp. are measured and used as four synchronization features (Ansari et al., 2015).

The authors would like to apologise for any inconvenience caused.

DOI of original article: http://dx.doi.org/10.1016/j.clinph.2016.06.018
* Corresponding author at: Department of Electrical Engineering (ESAT), KU Leuven, P.O. Box 2446, 3001 Leuven, Belgium.
E-mail addresses: amirhossein.ansari@kuleuven.be (A.H. Ansari), perumpij@mcmaster.ca (P.J. Cherian), anneleen.dereymaeker@uzleuven.be (A. Dereymaeker), Vmatic@singidunum.ac.rs (V. Matic), katrien.jansen@uzleuven.be (K. Jansen), a.dewispelaere@erasmusmc.nl (L. De Wispelaere), charlotte.dielman@zna.be (C. Dielman), jan.vervisch@uzleuven.be (J. Vervisch), r.swarte@erasmusmc.nl (R.M. Swarte), govaert@icloud.com (P. Govaert), gunnar.naulaers@uzleuven.be (G. Naulaers), maarten.devos@eng.ox.ac.uk (M. De Vos), Sabine.VanHuffel@esat.kuleuven.be (S. Van Huffel).

http://dx.doi.org/10.1016/j.clinph.2016.09.009
1388-2457/© 2016 International Federation of Clinical Neurophysiology. Published by Elsevier Ireland Ltd. All rights reserved.