KU LEUVEN

STADIUS
Center for Dynamical Systems, Signal Processing and Data Analytics

Citation/Reference	Moeyersons J., Amoni M., Willems R., Van Huffel S., Varon V. (2019), **R-DECO: An open-source Matlab based graphical user interface for the detection and correction of R-peaks** PeerJ Computer Science
Archived version	Author manuscript: the content is identical to the content of the published paper, but without the final typesetting by the publisher
Published version	https://www.biorxiv.org/content/10.1101/560706v1
Journal homepage	https://peerj.com/computer-science/
Author contact	Jonathan.moeyersons@kuleuven.be 0479427028
Abstract	
IR	Na.

R-DECO: An open-source Matlab based graphical user interface for the detection and correction of R-peaks

Jonathan Moeyersons [Corresp., 1, 2], **Matthew Amoni** [3, 4], **Sabine Van Huffel** [1, 2], **Rik Willems** [3, 4], **Carolina Varon** [1, 2]

[1] STADIUS Center for Dynamical Systems, Signal Processing and Data Analytics, Department of Electrical Engineering (ESAT), Katholieke Universiteit Leuven, Leuven, Belgium

[2] Imec, Leuven, Belgium

[3] Department of Cardiovascular Sciences, Katholieke Universiteit Leuven, Leuven, Belgium

[4] Department of Cardiology, University Hospitals Leuven, Leuven, Belgium

Corresponding Author: Jonathan Moeyersons
Email address: Jonathan.Moeyersons@esat.kuleuven.be

Many of the existing ECG toolboxes focus on the derivation of heart rate variability features from RR- intervals. By doing so, they assume correct detection of the QRS-complexes. However, it is highly likely that not all detections are correct. Therefore, it is recommended to visualize the actual R-peak positions in the ECG signal and allow manual adaptations.

In this paper we present R-DECO, an easy-to-use graphical user interface for the detection and correction of R-peaks. Within R-DECO, the R-peaks are detected by an adaptation of the Pan-Tompkins algorithm. Instead of using all the pre-processing steps of the latter algorithm, the proposed algorithm uses only an envelope-based procedure to flatten the ECG and enhance the QRS-complexes. The algorithm obtained an overall sensitivity of 99.60% and positive predictive value of 99.69% on the MIT/BIH arrhythmia database. Additionally, R-DECO includes support for several input data formats for ECG signals, three basic filters, the possibility to load other R-peak locations and intuitive methods to correct ectopic, wrong, or missed heartbeats. All functionalities can be accessed via the graphical user interface and the analysis results can be exported as Matlab or Excel files. The software is publicly available.

Through its easy-to-use-graphical user interface, R-DECO allows both clinicians and researchers to use all functionalities, without previous knowledge.

R-DECO: An open-source Matlab based graphical user interface for the detection and correction of R-peaks

Jonathan Moeyersons[1,2], Matthew Amoni[3,4], Sabine Van Huffel[1,2], Rik Willems[3,4], and Carolina Varon[1,2]

[1]**STADIUS Center for Dynamical Systems, Signal Processing and Data Analytics, Department of Electrical Engineering (ESAT), KU Leuven, Leuven, Belgium**
[2]**Imec, Leuven, Belgium**
[3]**Department of Cardiovascular Sciences, UZ Leuven, Leuven, Belgium**
[4]**University Hospitals Leuven, Department of Cardiology, Leuven, Belgium**

Corresponding author:
Jonathan Moeyersons[1]

Email address: jonathan.moeyersons@kuleuven.be

ABSTRACT

Many of the existing ECG toolboxes focus on the derivation of heart rate variability features from RR-intervals. By doing so, they assume correct detection of the QRS-complexes. However, it is highly likely that not all detections are correct. Therefore, it is recommended to visualize the actual R-peak positions in the ECG signal and allow manual adaptations.
In this paper we present R-DECO, an easy-to-use graphical user interface for the detection and correction of R-peaks. Within R-DECO, the R-peaks are detected by an adaptation of the Pan-Tompkins algorithm. Instead of using all the pre-processing steps of the latter algorithm, the proposed algorithm uses only an envelope-based procedure to flatten the ECG and enhance the QRS-complexes. The algorithm obtained an overall sensitivity of 99.60% and positive predictive value of 99.69% on the MIT/BIH arrhythmia database.
Additionally, R-DECO includes support for several input data formats for ECG signals, three basic filters, the possibility to load other R-peak locations and intuitive methods to correct ectopic, wrong, or missed heartbeats. All functionalities can be accessed via the graphical user interface and the analysis results can be exported as Matlab or Excel files. The software is publicly available.
Through its easy-to-use-graphical user interface, R-DECO allows both clinicians and researchers to use all functionalities, without previous knowledge.

INTRODUCTION

The electrocardiogram (ECG) is one of the primary screening and diagnostic tools of the cardiologist. It records the electrical activity of the heart, which generates the myocardial contractions. A crucial step in the study of the ECG is the location of the QRS-complexes. These complexes are the most prominent waveforms in the ECG, and contain an enormous amount of information about the state of the heart. This is why the detection of the QRS-complexes constitutes the basis for almost all automated ECG analysis algorithms (Kohler et al., 2002). Once these have been identified, more elaborated analyses can be performed, such as heart rate variability (HRV).

Four decades of automated QRS detection research has resulted in a variety of methods using different approaches. These methods can be stratified based on derivatives, digital filters, wavelet-transforms, classifiers, etc (Pan and Tompkins, 1985; Dohare et al., 2014; Fujii et al., 2013; Sharma and Sunkaria, 2016; Chen et al., 2006). Despite the wide methodological variety, most of these QRS detectors have the same algorithmic structure. This can be divided in two steps: pre-processing and decision making (Kohler et al., 2002).

In the pre-processing step the QRS-complex is highlighted and the other signal components are

suppressed to facilitate the detection. The resulting signal is then used to detect the occurrence of QRS-complexes in the decision making step. This is done by using either fixed or adaptive thresholds. Despite high detection rates, some QRS-complexes remain undetected. Reasons for this might be small amplitudes, wide complexes or contamination by noise (Arzeno et al., 2008). Therefore, in many algorithms an extra post-processing step is added for the exact determination of the temporal location of the detected QRS-complex.

One of the most established QRS detection algorithms is the Pan-Tompkins algorithm (Pan and Tompkins, 1985). Although it was developed in the eighties, it has proven to outperform many more elaborate algorithms. In this paper, an envelope-based procedure that enhances the QRS-complexes and flattens the rest of the ECG is used in combination with an adapted version of the threshold-based approach of the Pan-Tompkins algorithm. This method, which was proposed by our group in (Varon et al., 2015), combines the simplicity of an envelope-based procedure, while maintaining the accuracy of many more elaborate methods.

In a review paper, Elgendi et al. have compared the results of 22 beat detection algorithms on the MIT-BIH arrhythmia database (Elgendi et al., 2014). When comparing the results of the automated algorithms with expert annotations, they have shown that many algorithms obtained excellent accuracy. However, none of the algorithms reached perfection. This means that, no matter how good the QRS detection algorithm is, it is highly likely that not all annotations are correct. Therefore, it is recommended to visually inspect and review each signal before further analysis (Pichot et al., 2016).

Many of the existing ECG toolboxes have focussed on the derivation of HRV-analysis parameters from RR-intervals. This makes sense, since most of the available hardware include some kind of QRS-complex detection algorithm. However, this does not necessarily mean that the output of these devices are the raw RR-intervals. Many of these devices have a built-in post-processing algorithm, which compensates for false detections by averaging over a certain range of RR-intervals (Niskanen et al., 2004; Pichot et al., 2016; Vicente et al., 2013). However, for some analyses, such as ECG Derived Respiration (EDR) or Beat-to-beat Variability of Repolarization (BVR), it is of utmost importance that the actual R-peak of the QRS-complex is detected. Therefore, it is necessary to visualize the actual R-peak positions in the ECG signal and allow the possibility to make manual adaptations.

In this paper, we present R-DECO, a Matlab based, graphical user interface (GUI) for the detection and correction of R-peaks. This user interface includes the developed R-peak detection algorithm and provides the user with the possibility to correct possible false detections in a very straightforward way. R-DECO was developed by the biomedical data processing research team (BIOMED) at the Department of Electrical Engineering (ESAT) of KU Leuven. The software is freely available for Windows operating systems at [1].

The objective of this paper is to provide a detailed description of R-DECO, including the proposed R-peak detection algorithm and an overview of the different possibilities of this new software.

COMPUTATIONAL METHODS

R-peak detection

We developed an R-peak detection algorithm that is based on an enveloping procedure. It achieved a 99.60% sensitivity and 99.69% positive predictive value on the MIT/BIH Arrhythmia Database (Moody and Mark, 2001). The algorithm can be divided in three steps: pre-processing, decision and post-processing.

Pre-processing

In the presented approach, the filtering, derivative, squaring and integration steps of the Pan-Tompkins algorithm are replaced by one flattening step, which enhances the QRS-complexes and flattens the rest of the ECG (Varon et al., 2015). A visual explanation of the method is shown in Fig 1.

First, the upper (U) and lower (L) envelopes are computed from the ECG signal by the secant method. This method selects the segment with the steepest positive and negative slope in a user-defined window with length t. Once U and L are obtained, they are used to derive a flattened version of the ECG signal (F):

[1] http://homes.esat.kuleuven.be/jmoeyers/R_DECO.zip



Figure 1. Flattening procedure. The flattened ECG (F) is constructed by subtracting the lower envelope (L) from the upper envelope (U). This enhances the QRS-complex and flattens the rest of the ECG signal.

$$F = U - L \tag{1}$$

Since L is subtracted from U, the baseline is eliminated and only a positive signal, F, remains.

Decision

The locations of the QRS-complexes are found by detecting the peaks in the flattened ECG. These peaks are detected in three stages. First, all samples with an amplitude lower than the amplitude of the sample 80 ms further are selected. The 80 ms step size was experimentally defined. This results in the selection of the upward slopes. As a second step, only the upward slopes that are longer than the step size are selected in order to exclude small peaks. Finally, the maximum is selected in a window, with a length equal to the step size, that starts from the last selected sample of the upward slope.

On this selection of peaks, the adaptive thresholding procedure of the Pan-Tompkins algorithm is applied to define the peaks that correspond to the QRS-complexes.

Post-processing

The Pan-Tompkins algorithm generally produces satisfactory results in the detection of the QRS-complexes. However, some of the automatically generated RR-intervals might be physiologically unreasonable and need to be removed for further analysis. A slightly modified version of the search-back procedure as proposed in (de Chazal et al., 2003) was used for this purpose.

Once the positions of the QRS-complexes are identified, the original ECG is used to find the exact location of the R-peaks. The search for an R-peak is performed up to 50 ms from the peak in the flattened signal. This extra search is necessary, because the presence of large S-waves might shift the peak in the flattened signal towards the valley of the S-wave.

Evaluation on the Physionet MIT/BIH arrhythmia database

We used the MIT/BIH arrhythmia database to evaluate the proposed algorithm (Moody and Mark, 2001). This dataset consists of 48 half-hour ECG signals, which were recorded in the Boston's Beth Israel Hospital between 1975 and 1979. All recordings were annotated by two independent cardiologists who also made a distinction between normal and abnormal beats. In total, 110 122 heartbeats were annotated, of which 89 133 were labelled as normal. Each recording contains two channel ECG signals with a

sampling frequency of 360 Hz. In most records, one channel is lead II and the other channel is V1. However, we only used the first channel for the evaluation.

In Table 1 we listed the R-peak detection results of the proposed algorithm. With an envelope width of 300 ms and without post-processing, we obtained an overall sensitivity of 99.60% and positive predictive value (PPV) of 99.69%. With the post-processing, we obtain an overall sensitivity of 99.09% and PPV of 99.80%. These results are comparable with those in literature, especially with the Pan-Tompkins algorithm, which reaches a sensitivity of 99.76 % and a PPV of 99.56% (Elgendi et al., 2014).

Table 1. Performance of the R-peak detection algorithm on the Physionet MIT/BIH dataset.

Record	Total (beats)	TP (beats)	FP (beats)	FN (beats)	Se (%)	PPV (%)
100	2273	2273	0	0	100	100
101	1865	1864	4	1	99.95	99.79
102	2187	2187	0	0	100	100
103	2084	2084	0	0	100	100
104	2229	2227	3	2	99.91	99.87
105	2572	2542	41	30	98.83	98.41
106	2027	2023	2	4	99.80	99.90
107	2137	2130	0	7	99.67	100
108	1763	1739	80	24	98.64	95.60
109	2532	2532	0	0	100	100
111	2124	2123	2	1	99.95	99.91
112	2539	2539	0	0	100	100
113	1795	1795	0	0	100	100
114	1879	1877	6	2	99.89	99.68
115	1953	1953	0	0	100	100
116	2412	2388	2	24	99	99.92
117	1535	1535	0	0	100	100
118	2278	2278	1	0	100	99.96
119	1987	1987	2	0	100	99.90
121	1863	1861	12	2	99.89	99.36
122	2476	2476	0	0	100	100
123	1518	1518	0	0	100	100
124	1619	1619	0	0	100	100
200	2601	2595	8	6	99.77	99.69
201	1963	1958	0	5	99.75	100
202	2136	2120	14	16	99.25	99.34
203	2980	2749	20	231	92.25	99.28
205	2656	2641	2	15	99.44	99.92
207	1860	1855	8	5	99.73	99.58
208	2955	2941	2	14	99.53	99.93
209	3005	3005	0	0	100	100
210	2650	2582	3	68	97.43	99.88
212	2748	2748	1	0	100	99.96
213	3251	3250	0	1	99.97	100
214	2262	2259	3	3	99.87	99.87
215	3363	3354	0	9	99.73	100
217	2208	2202	0	6	99.73	100
219	2154	2154	1	0	100	99.95
220	2048	2047	0	1	99.95	100
221	2427	2425	2	2	99.92	99.92
222	2483	2475	21	8	99.68	99.16
223	2605	2605	0	0	100	100

Continued on next page

Table 1 – continued from previous page

Record	Total (beats)	TP (beats)	FP (beats)	FN (beats)	Se (%)	PPV (%)
228	2053	2044	59	9	99.56	97.19
230	2256	2256	0	0	100	100
231	1571	1571	0	0	100	100
232	1780	1780	17	0	100	99.05
233	3079	3070	0	9	99.71	100
234	2753	2753	1	0	100	99.96
Total	109 494	108 989	317	505	99.60	99.69

While these results are very promising, we can also observe that for some recordings only moderate detection results are obtained. This decrease in performance is generally due to poor signal quality, unusual morphology or stretches of extremely irregular rhythms. For instance, recording 116 and 208 contain stretches where the signal is lost in the first channel. However, the recordings with the highest amount of false detections are 108, 203 and 210. Correctly detecting the R-peaks in recording 108 has been proven difficult for many algorithms (Pan and Tompkins, 1985). It contains a lot of baseline wander and additionally, very tall and sharp P-waves. These characteristics make it difficult to distinguish P-waves from R-peaks and thus result in a high false positive count. However, the highest amount of false positives is observed in recording 203 (92.25% sensitivity). This might be explained by the extremely high percentage of PVC's present in the recording, almost 15%. Since the envelope width was fixed during the detection process, one may assume that the performance could be improved if manual adjustments were permitted. This holds as well for other records with PVC's, such as record 210.

From the analysis of the results, we could deduce two main factors that influence the results of the algorithm: (1) the envelope width and (2) the RR-post processing. The number of samples in the envelope is important, since it can be regarded as a filter of the RR-intervals. Smaller envelope widths might result in the enhancement of more peaks than only the R-peaks. This might be beneficial in the case of small R-peaks, but might also enhance artefact peaks. Larger envelope widths might cause adjacent R-peaks to be merged in the flattened signal. In practice, this might result in the failure of detecting premature heartbeats, which appear shortly after the previous heartbeat. In summary, a larger envelope width results in less false positives and more false negatives and the opposite is true for a small envelope width. A similar effect can be observed when the RR-intervals are post-processed. This increases the certainty of detection of the algorithm and thus results in less false positives. The downside is that, in the presence of abnormal rhythms, it also results in more false negatives.

SOFTWARE DESCRIPTION

The algorithms have been implemented with MATLAB R2018a. We used GUIDE, Matlab's GUI development environment, to design the GUI of R-DECO. The current subsection describes the possible input data formats and the user interface.

Input data formats

The standard input of the toolbox is raw or filtered ECG data. This can be both single- or multichannel ECG. Since a plethora of open formats exist for storing the ECG, it would be impossible to write supporting software for all formats (Niskanen et al., 2004). Therefore, we focussed on the data formats that are most commonly used by our clinical partners in the cardiology department of the UZ Leuven, Belgium. The following file formats are supported:

- ISHNE-Holter files (*.ecg)
- Matlab files (*.mat)
- European Data Format (*.edf)
- Text files (*.txt)

- Excel files (*.xls or *.csv)

An ISHNE-Holter file is organized in a header record, followed by a data block that contains all digital ECG samples. This file format was developed to facilitate data exchange and research in the field of Holter (Badilini, 1998). The software automatically extracts all ECG channels and also the sampling frequency.

A Matlab formatted file can contain one variable, up to an entire workspace. Therefore, if the file contains more than one variable, the user is prompted to select the variable containing the ECG signal. In the specific case that the selected file is a structure, the software allows to search within the structure until the ECG signal is selected.

A standard European Data Format, EDF, file consists of a header record and the data records (Kemp et al., 1992). It was originally intended for the digital storage and exchange of EEG and polysomnogram recordings, but currently it can store a variety of annotations and signals, such as EMG, ECG and many more (Kemp and Olivan, 2003). Since not all EDF files have the same standard labels, the user is prompted to identify the ECG channel(s). Additionally, the software attempts to identify the sampling frequency of the selected signal by scanning the file.

As an extra feature, the software allows the user to load a session. When the current session is interrupted, the session can be saved as a Matlab file. It includes all the analysis parameters, the ECG signal and the RR-intervals, if computed. When a previous session is loaded, the software restores the entire user interface to the moment on which the session was saved. This allows the user to pause and continue, whenever wanted.

User interface

The strength of this toolbox is that everything is operated through a single GUI. As shown in Fig 2, it can be divided in five segments: Data, Filter, Analysis Period, R-peak Detection and R-peak Correction. All segments are described below.



Figure 2. The graphical user interface of R-DECO. It can be divided in five segments: 1) Data, 2) Filter, 3) Analysis Period, 4) R-peak Detection and 5) R-peak Correction

Data

In the data panel, the user has the option to load data in two different ways: from a file or from the Matlab workspace. Both can be accessed via the respective pushbuttons.

When a file is selected, the software visualizes a small segment of the ECG signal. If the signal is inverted, the user can indicate this and the file will be inverted before further analysis. The software also scans the selected file for the sampling frequency. If this is not found, the software prompts the user to manually indicate the sampling frequency.

Finally, the data panel also contains a reset button. This button allows the user to reset the entire GUI. It empties all plots, restores all default variables and deletes all results. If the user has not yet saved the current analysis, the user is prompted to confirm the reset action to prevent unwanted loss of information.

Filter

Since ECG signals can be contaminated with noise, filtering is often essential for further analysis. R-DECO provides three basic filters: high pass, low pass and a notch filter.

The high pass filter consists of a zero phase, second order Butterworth filter. The low pass filter consists of a zero phase, fourth order Butterworth filter. Finally, also a zero phase notch filter is included. The latter could be used to remove the power-line interference. Important to note is that filtering actions are always executed on the original signal to ensure repeatability.

In order to aid the user in the selection of appropriate frequency threshold(s), R-DECO is able to compute and display the power spectrum. An estimate of the power spectrum is computed using the Welch method (Welch, 1967). As a default, we used a window of 500 samples with 60% overlap.

To visualize the effect of the filtering in the frequency domain, R-DECO displays both the filtered and the original power spectrum. Furthermore, the effect of the filtering in the time domain can also be investigated by checking the “Show Original” checkbox. This will overlay the original signal on top of the filtered signal (Fig. 3).



Figure 3. Example of a high pass filter. By clicking the ”Show Original” checkbox, the original signal (light blue) is overlaid on the filtered signal (blue).

Analysis period

In the Analysis Period panel the user has the possibility to define an analysis window. After pushing the “Define analysis period” button, the user has to select a window by clicking, dragging and releasing the mouse. The window is shown as a transparent patch over the data and can be enlarged, shrinked or moved with the mouse. After the initial window is drawn, the user can further modify the analysis window by changing the start time and/or the duration of the window.

When a window is selected, the user can accept the analysis window by pressing the ”Apply changes” button. This prompts the x-limits of the graph to match the analysis window and disables the window modifications. From here on, all further analysis will be performed solely on the selected window.

All the above is very useful when the to-be-selected time period is known in advance, however this is not always the case. Sometimes the selection of the analysis window is depending on certain patterns in the tachogram, hence the tachogram has to be constructed first. Therefore, if R-peaks have been detected already, the user is prompted to indicate whether he/she would like to keep the detected R-peaks.

R-peak Detection

The execution of the R-peak detection algorithm, as described in the first section, is initiated when the “Detect Peaks” button is pushed. However, before the actual algorithm is executed, the user is able to adjust the default parameters of the algorithm.

As can be seen from Fig 4, an epoch of 10 seconds of the ECG signal and its respective enveloped signal is shown as an example of the flattening procedure. The user can adjust the envelope size to the desired value and appreciate the changes by pressing the ”Apply” button. Additionally, an estimation of the average heart rate can be defined and the user has the possibility to automatically post-process the RR-intervals.



Figure 4. The R-peak parameter selection window. The user can adjust the envelope size, the average heart rate and indicate if RR post-processing is necessary. Pressing the Default button restores the default values.

Since some devices have built-in QRS detection algorithms and some researchers have their own preferred QRS detection algorithm, the software allows to load R-peak locations. These will be displayed the same way the R-peaks of the algorithm are displayed.

R-peak Correction

In case of heart rate variability studies, only the normal-to-normal RR-intervals need to be taken into account. This can be achieved by selecting the ectopic removal option. This option corrects ectopic beats, without altering the normal RR-intervals.

After finishing the detection process, either by detecting or loading the R-peaks, it is still possible that not all R-peaks are accurately detected. In R-DECO, the user can make manual and (semi-)automatic adjustments to the R-peak locations.

The manual methods are: add, adjust and delete. These methods can be activated by selecting the specific radiobuttons or via a context menu, which is linked to each R-peak. These manual methods allow the user to correct wrong or missing annotations either in all or in individual leads.

- Add: When this radiobutton is active, new R-peaks can be added by clicking in the ECG graph. The program selects the signal sample that is closest to the mouse position in a symmetric window of 300 ms around the mouse position. Upon mouse release, a new R-peak is added and the tachogram is adapted.
- Adjust: While hovering over the ECG graph, the R-peak closest to the mouse location is selected. After clicking on the desired R-peak, it can be moved by dragging the mouse. The movement of the selected R-peak is restricted by the previous and next R-peak. While adjusting the R-peak location, the tachogram is automatically updated. Upon mouse release, the new R-peak location is saved.
- Delete: While hovering over the ECG graph, the R-peak closest to the mouse location is selected. After clicking on the desired R-peak, more to-be-deleted R-peaks can be selected by dragging the mouse. Upon mouse release, the selected R-peaks are removed and the tachogram is adapted.

The three (semi-)automatic R-peak correction methods are: cross-correlation, peak conversion and maximum search.

- Cross correlation: For this method, a symmetrical window of 300 ms around each R-peak is selected. Then, all heartbeats are normalized by subtracting the mean and dividing it by the standard

deviation. Then, a trimmed average QRS-complex is computed of all the positive and 'negative' R-peaks. In this work a 'negative' R-peak is understood as the absence of an R-peak or the presence of a very prominent S-wave, also described as RS-complex.

The user is prompted to select either the positive or the 'negative' average heartbeat. If necessary, the user can also adjust the location of the R-peak on the selected template. This is all graphically displayed as shown in Fig. 5. Finally, the cross-correlation of every heartbeat is computed with the trimmed average and the R-peak is re-located, based on the highest correlation value.

- Peak conversion: The absolute amplitude of every R-peak annotation is compared with the previous and next sample's absolute amplitude. If it is bigger than the previous and smaller than the next, the location of the R-peak will be shifted forward, until an extremum is obtained. If it is the other way around, the location of the R-peak will be shifted backwards.



Figure 5. The template selection window. The user can select either the positive or 'negative' R-peak template and can shift the location of the R-peak if necessary.

This functionality avoids the 'jumping' of R-peaks from e.g. an actual R-peak to a pre-mature ventricular contraction, which might be the case when window search is applied. Furthermore, pressing the button more than once will not affect the relocation after one correction.

- Maximum search: Firstly, a symmetric window of 300 ms around each R-peak is selected. Secondly, the user is prompted to select either the maximum, minimum or absolute maximum. Based on this selection, the respective extremum within the window is selected as new R-peak location.

Save and export results

By default, the results are saved as a Matlab file. This file includes the R-peak locations and the RR-intervals, which can further be used for HRV-, BVR- or any other analysis that requires R-peak detection. Additionally, the software can export the results in two different ways: 1) an Excel file and 2) a Matlab file.

1. Excel file (*.xls): A new workbook is created with on the first sheet a general overview of the file: the number of channels, the sampling frequency, the duration of the signal and the duration of the analysis period. The number of additional sheets is defined by the number of channels, since for every channel, a separate sheet is created. This contains the R-peak locations, the RR-intervals and a number of basic metrics, such as the mean heart rate.

2. Matlab file (*.mat): This file contains a single structure named *data*. In accordance to the structure of the Excel file, a structure is created per channel. This contains the signal in the analysis window, the R-peak locations and the RR-intervals. This option is especially useful for further analysis in Matlab.

Data browser

In order to graphically correct the R-peaks it is important to have a clear view of the segment to be investigated. Therefore, after the R-peaks are detected, the software immediately displays the ECG signal with the R-peak annotations and the respective tachogram.

The window width of the x-axis can be adjusted in three different ways: 1) the range edit box, 2) the plus and minus buttons or 3) by using the zoom button. All three methods also adjust the width of the scroll bar.

The scroll bar can be used to slide through the signal. Since both axes are linked, both slide at the same time. The y-axis range in both axes is adjusted automatically according to the data within the selected range. However, if the “Fix y-limits” checkbox is selected, the range of the y-axis of the tachogram is fixed to the current limits. Since some users favour a tachogram that displays the RR-intervals, while others favour HR values, we made it possible to switch between the two.

An ECG recording with multiple channels might result in axes that become unclear. Therefore, R-DECO provides the user to switch view between different channels. This way the user can select one, multiple or all channels. If the channel labels are not present in the signal file, R-DECO names and numbers the channels itself: Channel 1, Channel 2, etc. However, it also provides the possibility to change these names. When the user clicks twice on a channel name in the listbox, a dialog box pops up that allows the user to choose a new name. This will be adjusted in the listbox, and also in the output files.

An extra feature is the possibility to take ’pictures’ of the axes. This saves both axes in a format of choice, without including any of the buttons or bars from the user interface. In order to design the axes to the user’s taste, R-DECO allows to change the line colors and the grid lines.



Figure 6. The data browsing options of R-DECO.

Preferences

The analysis settings of R-DECO can be adjusted via the Preferences menu. Note that the changes only apply for the current session and are not saved for the next session. This will be adjusted in a future release.

The Preference menu can be divided in four segments: Power Spectrum, Filter, Detection and Correction

- Power Spectrum: All variables of the Welch method can be adjusted here.

- Filter: The type and order of the filter can be defined here.
- Detection: The default input parameters for the R-peak detection algorithm can be adjusted here. Whenever the default button is pressed in the parameter selection window, see Fig 4, all parameters are reset to the values defined in this segment.
- Correction: For all correction methods, the user can define the window size in which the new R-peak is supposed to be located.

SAMPLE RUN

As a sample run, we used a 24 hour digital Holter signal that was recorded from a male subject with ischaemic heart disease. The idea was to investigate the temporal evolution in beat-to-beat variability of repolarization before spontaneous non-sustained ventricular tachycardia (nsVT). Before analysis could be performed, the nsVT episodes needed to be identified and the R-peaks needed to be detected.

Once the signal was loaded and the sampling frequency was defined, we first had a look at the power spectrum. According to the plot, no power line interference was present, since we could not observe a peak at 50 or 60 Hz. However, it was clear that most of the power was situated in the lower frequency bands. This could indicate the presence of baseline wander. Therefore, we high-pass filtered the signal with a cut-off frequency of 0.66 Hz. The result in the power spectrum can be observed in Fig. 7.



Figure 7. Result of the high pass filtering in the power spectrum.

Next, the nsVT episodes needed to be identified. However, the time stamps of the episodes were not known. Therefore, it was necessary to detect the R-peaks first. This way the nsVT episodes could be identified from the tachogram. An example of an nsVT episode, taken with the picture button, can be observed in Fig. 8



Figure 8. Example of an nsVT segment. The resulting tachogram is shown in the bottom plot.

Based on the example signal, we selected an envelope size of 300 ms, which provided the best results for this signal. This envelope size ensures the enhancement of the QRS-complexes, without skipping any beats. Additionally, we indicated that no post-processing of the RR-intervals is wanted, since we wanted to be able to detect nsVT segments as well.

The nsVT episodes were identified based on the resulting RR-intervals. From the start of one of these episodes, we selected 30 consecutive heartbeats (Thomsen et al., 2004). Only normal-to-normal intervals should be taken into account for BVR-analysis. Hence, (ventricular) ectopic and post-extrasystolic beats were removed for further analysis.

After the RR-intervals in the wanted analysis window were selected, the analysis results were saved. This was done by selecting "Save Results" on the menu bar and entering a file name. The results are then saved as a Matlab file and can be loaded for the following BVR-analysis. Note that you can also export the results as an Excel file.

POTENTIAL OF FUTURE GROWTH

R-DECO is the first step towards a complete ECG processing tool. At the moment, it focusses on accurate R-peak detection and intuitive correction options. Therefore, it is a complementary tool for existing HRV-analysis toolboxes, which tend to focus on the computation of HRV metrics from RR-intervals. Although some toolboxes already provide the possibility to detect R-peaks, their possibility to correct R-peak annotations is rather limited (Pichot et al., 2016; Rodenhauser et al., 2018; Vicente et al., 2013). Additionally, to the best of our knowledge, none of the existing toolboxes provide filtering, R-peak detection and correction all together.

The main advantage of R-DECO is the easy-to-use, intuitive GUI. All actions are performed in one window, which simplifies the use and reduces the learning time.

Several extra features are being developed and will be released in future versions. We intend to add support for other input file formats, such as Hierarchical Data Format 5 (HF) files, General Data Format (GDF) files, etc. However, most improvements will be in the amount of analysis options. Some of the first extra analysis options will be automatic signal quality detection, EDR and HRV-analysis.

CONCLUSION

R-DECO is a Matlab based GUI for detecting and correcting R-peaks in ECG signals. The goal of R-DECO is to provide a complete workflow from the raw signal to the tachogram. It includes an accurate R-peak detection algorithm, the performance of which is comparable to literature, and allows the user to graphically correct wrong or missing detections. Additionally, R-DECO supports a variety of ECG input file formats, which allows the processing of recordings directly from the recording device. This makes it a tool that can be used both by engineers, and clinicians.

We included some basic pre-processing options, such as three filters and the possibility to select an analysis window. The analysis results can be exported to the Matlab workspace or Excel for later analysis.

R-DECO is available free of charge upon request and can be downloaded from [2].

ACKNOWLEDGMENTS

Bijzonder Onderzoeksfonds KU Leuven (BOF): SPARKLE #: IDO-10-0358, The effect of perinatal stress on the later outcome in preterm babies #: C24/15/036, TARGID #: C32-16-00364; Agentschap Innoveren & Ondernemen (VLAIO): Project #: STW 150466 OSA + O&O HBC 2016 0184 eWatch; imec: SBO-2016, ICON: HBC.2016.0167 SeizeIT; European Research Council: The research leading to these results has received funding from the European Research Council under the European Union's Seventh Framework Programme (FP7/2007-2013) / ERC Advanced Grant: BIOTENSORS (n339804). This paper reflects only the authors' views and the Union is not liable for any use that may be made of the contained information. CV is a postdoctoral fellow of the Research Foundation-Flanders (FWO). RW is a Senior Clinical Investigator of the Research Foundation - Flanders (FWO).

REFERENCES

Arzeno, N. N., Deng, Z. D., and Poon, C. S. (2008). Analysis of first-derivative based qrs detection algorithms. *IEEE Transaction on Biomedical Engineering*, 55(2):478–484.

Badilini, F. (1998). The ishne holter standard output file format. *Annals of Noninvasive Electrocardiology*, 3(3):263–266.

[2] http://homes.esat.kuleuven.be/jmoeyers/R_DECO.zip

Chen, S.-W., Chen, H.-C., and Chan, H.-L. (2006). A real-time qrs detection method based on moving-averaging incorporating with wavelet denoising. *Computer Methods and Programs in Biomedicine*, 82(3):187–195.

de Chazal, P., Heneghan, C., Sheridan, E., Reilly, R., Nolan, P., and O'Malley, M. (2003). Automated processing of the single-lead electrocardiogram for the detection of obstructive sleep apnoea. *IEEE Transactions on Biomedical Engineering*, 50(6):686–696.

Dohare, A. K., Kumar, V., and Kumar, R. (2014). An efficient new method for the detection of qrs in electrocardiogram. *Computers and Electrical Engineering*, 40(5):1717–1730.

Elgendi, M., Eskofier, B., Dokos, S., and Abbott, D. (2014). Revisiting qrs detection methodologies for portable, wearable, battery-operated, and wireless ecg systems. *PLOS ONE*, 9(1):1–18.

Fujii, T., Nakano, M., Yamashita, K., Konishi, T., Izumi, S., Kawaguchi, H., and Yoshimoto, M. (2013). Noise-tolerant instantaneous heart rate and r-peak detection using short-term autocorrelation for wearable healthcare systems. *35th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, pages 7330–7333.

Kemp, B. and Olivan, J. (2003). European data format 'plus' (edf+), an edf alike standard format for the exchange of physiological data. *Clinical Neurophysiology*, 114(9):1755–1761.

Kemp, B., Värri, A., Rosa, A. C., Nielsen, K. D., and Gade, J. (1992). A simple format for exchange of digitized polygraphic recordings. *Electroencephalography and Clinical Neurophysiology*, 82(5):391–393.

Kohler, B. ., Hennig, C., and Orglmeister, R. (2002). The principles of software qrs detection. *IEEE Engineering in Medicine and Biology Magazine*, 21(1):42–57.

Moody, G. B. and Mark, R. G. (2001). The impact of the mit-bih arrhythmia database. *IEEE Engineering in Medicine and Biology Magazine*, 20(3):45–50.

Niskanen, J.-P., Tarvainen, M. P., Ranta-Aho, P. O., and Karjalainen, P. A. (2004). Software for advanced hrv analysis. *Computer Methods and Programs in Biomedicine*, 76(1):73–81.

Pan, J. and Tompkins, W. J. (1985). A real-time qrs detection algorithm. *IEEE Transaction on Biomedical Engineering*, 32(3):230–236.

Pichot, V., Roche, F., Celle, S., Barthélémy, J.-C., and Chouchou, F. (2016). Hrvanalysis: A free software for analyzing cardiac autonomic activity. *Frontiers in Physiology*, 7:557.

Rodenhauser, A., Good, W. W., Zenger, B., Tate, J., Aras, K., Burton, B., and MacLeod, R. S. (2018). Pfeifer: Preprocessing framework for electrograms intermittently fiducialized from experimental recordings. *Journal of Open Source Software*, 3(21):472.

Sharma, L. D. and Sunkaria, R. K. (2016). A robust qrs detection using novel pre-processing techniques and kurtosis based enhanced efficiency. *Measurement*, 87:194–204.

Thomsen, M. B., Verduyn, S. C., Stengl, M., Beekman, J. D., de Pater, G., van Opstal, J., Volders, P. G., and Vos, M. A. (2004). Increased short-term variability of repolarization predicts d-sotalol-induced torsades de pointes in dogs. *Circulation*, 110(16):2453–2459.

Varon, C., Caicedo, A., Testelmans, D., Buyse, B., and Huffel, S. V. (2015). A novel algorithm for the automatic detection of sleep apnea from single-lead ecg. *IEEE Transactions on Biomedical Engineering*, 62(9):2269–2278.

Vicente, J., Johannesen, L., Galeotti, L., and Strauss, D. G. (2013). Ecglab: User friendly ecg/vcg analysis tool for research environments. *Computing in Cardiology 2013*, pages 775–778.

Welch, P. (1967). The use of fast fourier transform for the estimation of power spectra: A method based on time averaging over short, modified periodograms. *IEEE Transactions on Audio and Electroacoustics*, 15(2):70–73.